

Date: 15-09-2022

GSR

X-ray Machine

The GSR for X-Ray machine as required by the MedAsk Radiology are as under:

a. Model Name/Number:	According to the unit
b. Brand:	Toshiba/Shimadzu/GE
c. Year of manufacture:	2011 & above
d. Generator:	High Voltage Generator (Inverter)
e. Table:	Floating top bucky table with Oscillating grid
f. X-ray System:	Diagnostic X-ray system
g. Tube Voltage:	50-150 KVA
h. Anode heat storage:	300 KHU
i. Tube Current:	500-630 mA (3 phase)
j. Nominal Current:	380v to 440v +- 10%, 50Hz
k. X-ray type:	Ceiling or Column mounted
l. X-ray tube Assembly:	Matches the X-ray Unit Serial no.
m. X-Ray tube support:	Matches the X-ray Unit Serial no.
n. DR/CR System:	DR/CR System as per availability
o. DR/CR Make:	Branded/Chinese as per availability
p. DR Plate size:	17 x 17
q. Printer:	Branded
r. Printer tray size:	As per specification
s. Services:	After Sale services required locally
t. Origin:	Refurbished

Note:

Please send your sealed proposal to MedAsk Pvt. Ltd, First floor, AWT Plaza, Main Mall Road Rawalpindi addressed to **Consultant Procurement (Brig. Dr. Riaz Ahmed Raja Retd.)**

**Consultant Procurement
Brig. Riaz Ahmed Raja (R)**



Date: 15-09-2022

GSR

Ultrasound Machine

1. The GSR for Ultrasound machine as required by the MedAsk Radiology are as under:

- | | |
|-------------------------|--|
| a. Model Name/Number: | As per machine |
| b. Brand: | GE/Aloka/Toshiba |
| c. Year of Manufacture: | 2011 and above |
| d. Probes Included: | Convex & linear |
| e. Optional Probes: | Cardiac & TVS |
| f. USG System: | Latest |
| g. Luminance: | 3D/4D Imaging suite |
| h. Flow: | Advanced dynamic flow |
| i. Connector: | 3-Active transducer sockets |
| j. Monitor: | high resolution 16/17/19 inch IPS monitor |
| k. Control panel: | Fully customizable, mode sensitive touch panel |
| l. Imaging Type: | Comprehensive on board reporting |
| m. Connectivity: | 3.0 connectivity |
| n. Data Management: | Export imaging using USB, network or
Built-in DVD drive |
| o. Origin: | Refurbished |

Note:

Please send your sealed proposal to MedAsk Pvt. Ltd, First floor, AWT Plaza, Main Mall Road Rawalpindi addressed to **Consultant Procurement (Brig. Dr. Riaz Ahmed Raja Retd.)**

**Consultant Procurement
Brig. Riaz Ahmed Raja (R)**